

Falcon® Pipet Controller




Instruction Manual

Catalog Number:
357469

FALCON®

A Corning Brand



1**2****3**

- A – Aspiration button – PP
- B – Dispense button – PP
- C – Suction speed switch – PP
- D – Dispense mode switch – PP
- E – Display
- F – Nose piece – PP
- G – Pipet holder – SI
- H – Membrane filter – PP/PTFE
- J – Connector gasket – SI
- K – Batteries cover – PP
- L – Battery – NiMH, AAA, 1.2V
- M – Charging stand – PP
- N – Charger 9V: EU, US, UK, AU
INPUT: 100-240V, 50/60Hz, 0.3A
OUTPUT: AC DC 9V, 230mA

Casing – PP

PP: Polypropylene
PTFE: Polytetrafluoroethylene
SI: Silicone






4**5****6****7****8**

Table of Contents

- 1.0 Introduction5
- 2.0 Work Safety Instructions5
- 3.0 Limitations of Use.....6
- 4.0 Turning on the Falcon® Pipet Controller6
- 5.0 Aspirating and Dispensing Liquids6
- 6.0 Troubleshooting7
- 7.0 Replacing the Filter.....8
- 8.0 Battery Charging Procedure.....8
- 9.0 Maintenance9
- 10.0 Components..... 10
- 11.0 Ordering Information 10
- 12.0 Spare Parts 10
- 13.0 Limited Warranty 11
- 14.0 Equipment Disposal..... 11

1.0 Introduction

The Falcon® Pipet Controller is a device intended for pipetting liquids with the use of measuring pipets. It can work with all types of glass or plastic pipets in the volume range from 0.5 mL to 100 mL.

Two dispense modes permit selection of dispensing intensity depending on the user's needs. The selected setting of the pipet controller mode is shown on the display (Figure 1E). Figure 1 shows the external parts of the pipet controller with a description of the materials used. Display indicators are shown in Figure 2.

2.0 Work Safety Instructions

Use of this product in any manner not specified by the manufacturer may impair the safety protection provided by the equipment and may result in physical damage, personal injury, and/or errors in pipetting of liquids. Please read all operating instructions in this manual prior to use of this equipment.



This symbol indicates a potential risk and alerts you to proceed with caution.

- Using the device inconsistently with the operating instructions may result in damaging the device.
- The device should be serviced only at an authorized service center, otherwise the manufacturer will be relieved from any liability under the warranty.
- Only original spare parts and accessories, recommended by the manufacturer, shall be used.
- Only the original charger, supplied by the manufacturer, shall be used for charging the batteries.
- In case of incorrect functioning of the pipet controller, work shall be stopped. The device shall be cleaned according to Section 9 and sent for repair to an authorized service center.
- In the case of mechanical damage to the casing, the device shall be immediately sent for repair to an authorized service center.
- The use of excessive force during work shall be avoided.



WARNING!

- While working with the pipet controller general safety regulations regarding risks related with laboratory work should be observed. Protective clothing, goggles, and gloves should be worn.
- The pipet controller shall be used only for measuring liquids in conditions specified by the manufacturer, which are limited due to the chemical and mechanical resistance of the device, as well as the user safety.
- The information and instructions provided by the manufacturers of the reagents must be observed.

3.0 Limitations of Use

- The pipet controller should not be used for measuring substances, the vapors of which damage the following plastics: PP, SI, EPDM, POM.
- The pipet controller should not be used in an environment where explosion risk is present.
- Flammable liquids should not be measured, in particular substances with flash-point below 0°C (ether, acetone).
- The pipet controller should not be used for drawing acids with a concentration above 1 mol/L.
- The pipet controller should not be used for drawing solutions with a temperature above 50°C.
- The pipet controller should be operated in the temperature range from +10°C to +35°C.
- The pipet controller is suitable for general laboratory use only. It must be used only by personnel who know the health risks associated with the substances that are normally used with this instrument.

4.0 Turning on the Falcon® Pipet Controller

The pipet controller is switched on by pressing any of the buttons (Figures 1A, B, C, D). The display will show the selected aspiration mode, dispense mode, and the battery level indicator (Figure 2.3). Examples of the display indications are shown in Figure 5. The battery is discharged and requires charging if the indicator displays one “bar” (Figure 5.3). The batteries need to be recharged if the indicator does not display “bars” (Figure 5.4). When the batteries are fully charged, the indicator displays three “bars” (Figure 5.1).

- The pipet controller will switch off automatically if not used for 5 minutes.
- The pipet controller should be charged only with the original charger.
- The main voltage should conform to the specification on the charger.
- Charging should be done in accordance with section 8 of this instruction for use.

5.0 Aspirating and Dispensing Liquids

Attaching a pipet

CAUTION: Before attaching a pipet, check to ensure the pipet is not damaged, has no dents or sharp edges in the gripping section. Check to ensure the gripping section is dry.

The pipet controller should be gripped as close to the upper end as possible and carefully inserted into the pipet controller holder until resistance is felt (Figure 3A). Do not apply excessive force, to prevent damage to thin pipets controller and to avoid injury risk. A pipet that has been correctly attached and sealed in the holder should not tilt to the side.

After attaching a pipet, hold the device in such a way as to keep the pipet controller in a vertical position with an attached pipet. When not in use, it is not recommended to leave the pipet attached for long periods, for example overnight or over a weekend.

NOTE: Do not rest the pipet controller on the bench if there is liquid in the pipet.

Filling the pipet

Before aspirating is started, set the suction speed switch using the HIGH/LOW switch (Figure 1C):

- HIGH position: fast aspirating (Figure 2.1)
- LOW position: slow aspirating (Figure 2.2)

It is recommended to choose the LOW position for pipets with a volume up to 5 mL, and the HIGH position for pipets with a volume greater than 5 mL.

Holding the pipet controller vertically, immerse the pipet end in the liquid to be drawn up (Figure 3B), and press the aspiration button gently. The pipet filling speed depends on the pressure applied to the aspirating button. The greater the pressure applied, the faster the liquid is aspirated into the pipet.

It is recommended to draw a slightly larger liquid volume than required (due to meniscus above the required volume mark), reducing the aspiration speed, particularly in the final filling stage, so as not to overfill the pipet.



Setting the Volume

After filling the pipet, the end should be dried with an absorbent paper that will not leave impurities, in order to remove any excess solution on the outside surface of the pipet. To set the required liquid volume precisely, press the dispense button gently (Figure 3C), dispense the excessive liquid from the pipet until the meniscus of the liquid aligns exactly with the required volume mark on the pipet.

Emptying the Pipet

Holding the vessel in an angled position, place the pipet end in contact with the vessel wall and press the dispense button gently (Figure 3C). The dispensing intensity may be adjusted depending on pressure applied to the dispense button. The greater the pressure applied, the faster the outflow of liquid from the pipet.

The pipet controller has two dispense modes. The dispense mode is selected with the BLOW/GRAV switch (Figure 1D).

- Gravity mode marked with the  icon on the display (Figure 2.4): Dispensing is effected in gravity mode, which means that the liquid flows out of the pipet by its own weight.
- Blow out mode marked with the  icon on the display (Figure 2.5): Dispensing is effected in gravity mode; however, when the dispense button is pressed to the middle position, the pump is started and fast emptying of the pipet with a blow out is affected.

CAUTION: During gravimetric dispensing the pipet may not be completely emptied due to the characteristics of pipets used with the pipet controller.

6.0 Troubleshooting

If during your work the pipet controller is not functioning correctly, check the cause and correct the fault.

Problem	Possible Cause	Action
The pipet falls out (the holding force of the pipet is too small), or tilts to the side too much.	The pipet holder (Figure 1G) is dirty or wet.	Take out the pipet holder, and clean, wash, and dry it.
	The pipet holder is damaged.	Replace the pipet holder with a new one.
The pump is working, but the pipet controller aid does not draw liquid or draws liquid very slowly.	The filter (Figure 1H) is dirty.	Take out the pipet holder, take out the filter; if it is dirty, replace it with a new one.
	The pipet holder and/or the connector gasket (Figure 1J) is damaged.	Replace the mechanically damaged elements with new ones.

Problem	Possible Cause	Action
Liquid leaks from the pipet (the aspiration and the dispense buttons are not pressed).	The pipet is damaged.	Check the pipet for damage (cracks, dents); if present, replace the pipet with a new one.
	The pipet is inserted incorrectly.	Check whether the pipet has been correctly inserted in the pipet holder.
	The pipet holder, the filter, or the connector gasket is installed incorrectly.	Check whether all parts are present and correctly installed.
	The pipet holder and/or the connector gasket is damaged (Figures 1G, 1J).	Replace the mechanically damaged elements with new ones.

If the above actions do not remedy the incorrect functioning of the pipet controller, the device should be sent to the nearest authorized service center.

Before being serviced, the pipet controller should be cleaned and decontaminated. Written details including the precise specification of solutions used and the type of laboratory in which the device was used, should be sent with the product.

7.0 Replacing the Filter

NOTE: The work safety instructions given in section 2 should be observed when disassembling the pipet controller.

The filter replacement is necessary if drawing efficiency deterioration is observed. The direct reason may be a dirty filter after a long period of use. The procedure is shown in (Figure 4).

- Remove the pipet.
- Unscrew the nose piece (Figure 4A).
- Remove the membrane filter (Figure 4A) and the pipet holder (Figure 4B).
- Rinse the holder using a wash bottle (Figure 4C).
- Blow liquid out of the holder and set aside until it is completely dry.
- Install the new membrane filter (Figure 4D), and assemble the device in reverse order.

8.0 Battery Charging Procedure

The pipet controller may be charged only with the original charger. The main voltage must conform with the specification on the charger.

NOTE: Using other chargers than the original one may damage the battery of the pipet controller.

Charging the battery in the pipet controller is controlled by a time circuit which controls the entire process. When the battery has been fully charged, the charging circuit disconnects automatically.

The pipet controller comes with 3 x AAA, NiHM batteries. In order to prolong the battery life span please consider the following:

- Before the pipet controller is activated for the first time the battery should be charged.
- Full battery charge allows for approximately 6 to 8 hours of continuous work.
- Allow the battery status to come down to 1 bar before charging the battery.
- Frequent and/or incomplete charging will shorten the battery life.
- Charge overnight to avoid process disruptions.
- When low battery level is indicated, you can continue to work while charging.
- Condition the battery every month. This is done by allowing the battery to discharge completely, then charging it up to full (charge overnight to avoid process disruptions).
- If the pipet controller will not be used for several weeks, charge the battery fully, remove it from the equipment, and store it separately.
- Never mix batteries of different manufacturers, capacity, type (i.e., NiCd, NiMH, Li-ion), or usage time (old with new).
- When needed, be sure to have a replacement set of batteries available.
- Never use a battery that is damaged or leaking.

Charging Conditions:

1. Charging temperature: 10°C to 35°C.
2. Charging the battery is carried out through a charger (power supply) by direct connection to the power source, or indirectly by means of a charging stand (Figure 1M). Battery charging is indicated by successive lighting of “bars” in the display order 4-3-2-1-4-3-2-1 (Figure 5).
3. Full charging time: 7 to 8 hours.
4. The batteries are charged when all 3 “bars” are displayed simultaneously (Figure 5.1).

9.0 Maintenance

Cleaning

The external parts of the pipet controller may be cleaned with a swab moistened with isopropyl alcohol.

The nose piece and the pipet holder may be autoclaved at 121°C for 20 minutes. After autoclaving, allow the silicone pipet holder to cool. The filters included in the set may be sterilized by autoclaving at 121°C for not more than 15 minutes.

Ultra violet (UV) sterilization

The outer body of the pipet controller is UV-resistant, however the recommended distance from the radiation source to the exposed element should not be less than 50 cm.

Prolonged and very intense UV exposure can cause decoloration of pipet controller parts, without affecting its performance.

Storage

The pipet controller should be stored in a dry place. Recommended storage temperature: -20°C to +50°C.

When not in use, the pipet controller should be placed in the charging stand included with the pipet controller (Figures 7A, 7B).

NOTE: Do not store the pipet controller with a filled pipet.

10.0 Components

The pipet controller set is supplied with the following components:

- Universal charger with set of adapters
- Charging stand
- Membrane filter 0.2 µm
- Instruction manual
- Quality control certificate

11.0 Ordering Information

The Falcon® Pipet Controller (Cat. No. 357469) comes with a universal charger and a set of adapters in different versions: EU, US, UK, and AU. Choose your country's adapter and connect to the housing.

To mount the adapter, it should be inserted into the slots of the housing (Figure 6N) in the direction of the arrow, until you hear a click.

To remove or change the adapter, simply press the "PUSH" button in the direction of the arrow, hold the button down, remove the adapter in the direction of the arrow.

12.0 Spare Parts

Item in Figure 1	Description	Cat. No.	Qty/Pk
F	Nose cone	357476	1
G	Silicone pipet holder	357474	1
H	PTFE filter 0.2 µm	357472	5
	PTFE filter 0.45 µm	357473	5
L	Battery set, 3 x AAA, NiHM	357486	3
N	Universal charger, 9V with set of adapters EU, US, UK, AU	357489	1
M	2-position charging stand	357490	1
K	Battery cover	357488	1
J	Connector gasket	357477	1

13.0 Limited Warranty

Corning Incorporated (Corning) warrants that this product will be free from defects in material and workmanship for a period of one (1) year from date of purchase. CORNING DISCLAIMS ALL OTHER WARRANTIES WHETHER EXPRESSED OR IMPLIED, INCLUDING ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR OF FITNESS FOR A PARTICULAR PURPOSE. Corning's sole obligation shall be to repair or replace, at its option, any product or part thereof that proves defective in material or workmanship within the warranty period, provided the purchaser notifies Corning of any such defect. Corning is not liable for any incidental or consequential damages, commercial loss or any other damages from the use of this product.

This warranty is valid only if the product is used for its intended purpose and within the guidelines specified in the supplied instruction manual. This warranty does not cover damage caused by accident, neglect, misuse, improper service, natural forces or other causes not arising from defects in original material or workmanship. This warranty does not cover batteries or damage to paint or finish. Claims for transit damage should be filed with the transportation carrier.

In the event this product fails within the specified period of time because of a defect in material or workmanship, contact Corning Customer Service at: USA/Canada 1.800.492.1110, outside the U.S. +1.978.442.2200, visit www.corning.com/lifesciences, or contact your local support office.

Corning's Customer Service team will help arrange local service where available or coordinate a return authorization number and shipping instructions. Products received without proper authorization will be returned. All items returned for service should be sent postage prepaid in the original packaging or other suitable carton, padded to avoid damage. Corning will not be responsible for damage incurred by improper packaging. Corning may elect for onsite service for larger equipment. Some states do not allow limitation on the length of implied warranties or the exclusion or limitation of incidental or consequential damages. This warranty gives you specific legal rights. You may have other rights which vary from state to state.

No individual may accept for, or on behalf of Corning, any other obligation of liability, or extend the period of this warranty.

For your reference, make a note of the serial and model number, date of purchase, and supplier here.

Serial No. _____ Date Purchased _____

Model No. _____

Supplier _____

14.0 Equipment Disposal



According to Directive 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE), this product is marked with the crossed-out wheeled bin and must not be disposed of with domestic waste.

Consequently, the buyer shall follow the instructions for reuse and recycling of waste electronic and electrical equipment (WEEE) provided with the products and available at www.corning.com/weee.

Warranty/Disclaimer: Unless otherwise specified, all products are for research use or general laboratory use only.* Not intended for use in diagnostic or therapeutic procedures. Not for use in humans. These products are not intended to mitigate the presence of microorganisms on surfaces or in the environment, where such organisms can be deleterious to humans or the environment. Corning Life Sciences makes no claims regarding the performance of these products for clinical or diagnostic applications. *For a listing of US medical devices, regulatory classifications or specific information on claims, visit www.corning.com/resources.

Corning's products are not specifically designed and tested for diagnostic testing. Many Corning products, though not specific for diagnostic testing, can be used in the workflow and preparation of the test at the customers discretion. Customers may use these products to support their claims. We cannot make any claims or statements that our products are approved for diagnostic testing either directly or indirectly. The customer is responsible for any testing, validation, and/or regulatory submissions that may be required to support the safety and efficacy of their intended application.

CORNING

Corning Incorporated
Life Sciences

www.corning.com/lifesciences

NORTH AMERICA

t 800.492.1110

t 978.442.2200

ASIA/PACIFIC

Australia/New Zealand

t 61 427286832

Chinese Mainland

t 86 21 3338 4338

India

t 91 124 4604000

Japan

t 81 3-3586 1996

Korea

t 82 2-796-9500

Singapore

t 65 6572-9740

Taiwan

t 886 2-2716-0338

EUROPE

CSEurope@corning.com

France

t 0800 916 882

Germany

t 0800 101 1153

The Netherlands

t 020 655 79 28

United Kingdom

t 0800 376 8660

All Other European Countries

t +31 (0) 206 59 60 51

LATIN AMERICA

grupoLA@corning.com

Brazil

t 55 (11) 3089-7400

Mexico

t (52-81) 8158-8400